

# **EC Declaration of Conformity**

Manufacturer:

Name: Beijing Ruicheng Medical Supplies Co., Ltd.

Address: No. 558 Zhangzikou, Yangsong Town, Address: Koningin Julianaplein 10, 1e Verd,

Huairou District, 101400 Beijing, China

Whose single Authorized Representative:

Name: Lotus NL B.V.

2595AA, The Hague, Netherlands.

We, the manufacturer, herewith declare that the product

## **Disposable Lancets**

Models: 21G I, 23G I, 25G I, 26G I, 27G I, 28G I, 30G I, 33G I; 21G II, 23G II, 25G II, 26G II, 27G II, 28G II, 30G II, 33G II; 21GIV, 23GIV, 25GIV, 26GIV, 27GIV, 28GIV, 30GIV, 33GIV; 21GV, 23GV, 25GV, 26GV, 27GV, 28GV, 30GV, 33GV; 21GM3, 23GM3, 25GM3, 26GM3, 27GM3, 28GM3, 30GM3, 33GM3; 21GM4, 23GM4, 25GM4, 26GM4, 27GM4, 28GM4, 30GM4, 33GM4;

### UMDNS-Code: 10440

Meet the provisions of Directive 93/42/EEC which apply to them.

The medical device has been assigned to Class IIa according to Annex IX of the Directive 93/42/EEC. It bears the mark

The product concerned has been manufactured under a quality management system according to Annex V of Directive 93/42/EEC.

Compliance of the designated product with the Directive 93/42/EEC has been assessed and certified by the Notified Body

## TÜV Rheinland LGA Products GmbH Tillystraße 2, 90431, Nürnberg, Germany

Certificate No.: DD 60141252 0001

Issue date: 2019-08-05 Expiry date: 2024-05-27

Following the procedure relating to the EC Declaration of Conformity set out in annex II of directive 93/42/EEC. This Declaration of Conformity covers all medical devices as specified in the product list belonging to this declaration and is only valid in connection with a batch specific Certificate of Compliance for all products concerned bearing the CE mark.

The above mentioned declaration of conformity is exclusively under the responsibility of

Company: Beijing Ruicheng Medical Supplies Co., Ltd.

Address: No. 558 Zhangzikou, Yangsong Town, Huairou District, 101400 Beijing, China

Beijing July 5, 2019.

Name! Yuechao Li General Manager

Legally binding signature, function